

PRODUCT: Niflex35 Nitrile Gloves (Blue)

ITEM #: H1012K

DESCRIPTION

- 3.5mil Textured Grip
- Powder Free Glove
- Economical puncture resistance with great comfort, dexterity, and tactile sensitivity
- Textured for better grip in wet or dry environments
- Contains no natural rubber latex proteins
- Ambidextrous
- Accelerator Free - formulated without chemical accelerator and reduces irritants to the skin.
- CPF certified - gloves have been tested and clean certified from common pathogens.
- Made from component materials in compliance with Federal Regulations for food contact 21CFR, 170-199. CFIA Approved.

Test	Object		Specification
Dimension(mm)	Palm Width	Extra Small	75mm ± 5
		Small	85mm ± 5
		Medium	95mm ± 5
		Large	105mm ± 5
		Extra Large	115mm ± 5
	Length		Min. 240mm
	Thickness (Single wall)	Cuff	Min. 0.06mm
Palm		Min. 0.07mm	
Finger		Min. 0.07mm	
Physical Properties	Tensile Strength (Mpa)	Before Aging	Min. 14
		After Aging	Min. 14
	Elongation (%)	Before Aging	Min. 500
		After Aging	Min. 400



Accelerator Free: These gloves are “Accelerator Free” which means allergic reaction would only affect a very, very small group of hypersensitive people. The accelerator chemicals (from the past) are known to cause type IV hypersensitivity (contact dermatitis).

Using a different form of cross-linker (which replace the accelerators and Sulphur) removes a huge component that cause the type IV reactions.

TESTING STANDARD: Gloves dimension and physical properties meet ASTM D6319 specification

ISO 374-2	FDA 21 CFR 177 .2600	ISO 21420	EN 1186-1	Food Approved
✓	✓	✓	✓	✓

CPF: These “Certified Pathogen Free” gloves have been tested and clean certified from accidental pathogen contamination during the manufacturing and production process. Gloves have been tested and clean certified for the following pathogens: Staphylococcus aureus, Escherichia coli, Listeria, Salmonella.

COMPLIANCE: Gloves are made from component materials that comply with FDA Regulation for food contact FDA 21 CFR 177.2600 and EU food contact text for overall migration BS EN 1186-1:2002. Components used in making Niflex gloves may cause allergic reactions with some people. Niflex gloves are latex free.

EN ISO 21420 Clause 7.2.1.1 (c)

EU declaration of conformity is accessible at www.amerisan.com

PPE Regulation (EU) 2016/425, Annex II, 1.4 (I)

In compliance with Personal Protective Equipment Regulation (EU) 2016/425

PPE Regulation (EU) 2016/425, Annex II, Clause 1.4(i)

Hand protection against minimal risk only.

EN ISO 21420, Clause 7.2.2

EN ISO 21420, Clause 7.3.4

EN ISO 21420, Clause 7.3.11

Before usage, inspect the gloves for any defect or imperfections.

EN ISO 374-1:2016/Amd.1:2018, Clause 7

Check that glove does not present holes, cracks. Tears. Or color change and discard any gloves presenting such.

EN ISO 21420, Clause 7.3.12

Where relevant, a list of the substances contained in the glove which are known to cause allergies, per listed in Annex G of EN ISO 21420:2020, shall be supplied on request.

Donning – Hold glove by the bead with one hand. Align the glove thumb with your other hand thumb and slide your hand into the glove, one finger into each glove finger. Pull by the glove palm to get a good fit. Don the other glove by the same procedure.

Doffing – Hold glove bead and pull toward the finger until the glove come off.

PACKAGING: 100 pieces/dispenser, 10 dispensers/carton for size S-XL, 90 pieces/dispenser, 10 dispensers/carton for size XXL.

HANDLING AND STORAGE: Store in a dry place at room temperature. Exposed product should be shielded from direct sunlight, intense artificial light, X-ray machines and other sources of ozone. UMBO[®] Niflex nitrile gloves have a shelf life of five years from the date of manufacture under ideal storage conditions.



Company name: UMBO® LLC
 Address: PO Box 228 Gradyville, PA 19039
 Website: umbousa.com

Common Pathogen Free Certified

These gloves have been tested and clean certified for the following pathogens:
Staphylococcus aureus, Escherichia coli, Listeria, Salmonella.

Test Parameter	Unit	Powder free nitrile examination gloves, blue, non sterile, Low Derm	Test Method
		BBR01890	
Staphylococcus aureus	CFU/50cm ²	< 1 (ND)	AS 1766.2.4-1994 (spread plate)
Escherichia coli	CFU/50cm ²	< 1 (ND)	AS 1766.2.3-1992
Listeria	-	Absent	AOAC 997.03
Salmonella	Present/Absent	Absent	ISO 18593 and ISO 6579-1

Our third-party testing by BP Food Environmental Testing verifies that UMBO® Niflex® Nitrile gloves carrying the Common Pathogen Free Certified badge are free from accidental pathogen contamination during the manufacturing and production process* of our Niflex® Nitrile single use gloves.

The Common Pathogen Free verified badge certifies that all our UMBO® Niflex® Nitrile Gloves have been tested and passed our strict safety standards providing you the highest level of protection and safety.



* The hidden risks of glove contamination result from the largely unregulated production and manufacturing of single-use gloves, with limited and infrequent glove compliance testing on import to the U.S.